

ICP-MS Method Reporting Limit

		µg/L			µg/L
Lithium	Li	0.3	Nickel	Ni	0.056
Beryllium	Be	0.06	Copper	Cu	0.041
Boron	B	2	Zinc	Zn	0.7
Sodium	Na	40	Arsenic	As	0.4
Magnesium	Mg	5	Selenium	Se	2.3
Aluminum	Al	5	Selenium	Se	1.5
Silica	Si	10	Selenium	Se	0.6
Phosphorus	P	50	Strontium	Sr	0.029
Potassium	K	10	Molybdenum	Mo	0.057
Calcium	Ca	10	Silver	Ag	0.013
Calcium	Ca	10	Cadmium	Cd	0.041
Vanadium	V	0.2	Tin	Sn	5
Chromium	Cr	0.11	Antimony	Sb	0.04
Manganese	Mn	0.05	Barium	Ba	0.2
Iron	Fe	10	Thallium	Tl	0.04
Cobalt	Co	0.028	Lead	Pb	0.06